

Illuminance Sensitivity Factors for OL 730C

Configuration	Minimum	Maximum
730-5-PF with		
730-5A	7×10^{-7} lux	1.3×10^5 lux
730-5H	7×10^{-8} lux	1.3×10^5 lux
740-15	7×10^{-10} lux	$\sim 7 \times 10^{-3}$ lux
730-5-PF & 85-O with		
730-5A	2.5×10^{-5} lux	4.5×10^6 lux
730-5H	2.5×10^{-6} lux	4.5×10^6 lux
740-15	2.5×10^{-8} lux	$\sim 2.5 \times 10^{-1}$ lux
730-5-PF & IS-670 with		
730-5A	8.5×10^{-5} lux	1.5×10^7 lux
730-5H	8.5×10^{-6} lux	1.5×10^7 lux
740-15	8.5×10^{-8} lux	$\sim 8.5 \times 10^{-1}$ lux
730-5-PF & 730-7G with		
730-5A	2×10^{-5} lux	3.7×10^6 lux
730-5H	2×10^{-6} lux	3.7×10^6 lux
740-15	2×10^{-8} lux	$\sim 2 \times 10^{-1}$ lux
730-5-PF & 730-7G & 85-O		
730-5A	5×10^{-4} lux	9×10^8 lux
730-5H	5×10^{-5} lux	9×10^8 lux
740-15	5×10^{-7} lux	~ 5 lux



OPTRONIC LABORATORIES

A Gooch & Housego Company

Optronics Laboratories, Inc. 4632 36th Street, Orlando, FL 32811
 Tel: 1 407 422 3171 Fax: 1 407 648 5412 Email: info@olinet.com

Information Sheet (IS15) / Rev. 9-94